

Criteria for rejection of Environmental Lead samples

After following recommended collection procedures for dust, paint, and soil and submission of samples with required requisition form (DCH-0558 March 2006) the lab will reject samples according to the following criteria:

1. Insufficient amount, < 1.5g for soil or 300mg for paint
2. Mixed or unfit sample, containing unidentified debris or matrix
3. Unsuitable container, provided graduated 50ml cent. tube recommended
4. Container not numbered or labeled properly (double identifier)
5. Container identification does not match requisition
6. Sample type/matrix not acceptable for approved methods

Note that final report will state NOT ANALYZED and the rejected criteria above will appear as a comment on that sample.